


Roping in uncertainty – measuring the tensile strength of steel wire ropes

Riaan Bergh
27 September 2016

Presentation outline



The outcome



Why test?



The test environment

A screenshot of a software application showing a table of measurement results. The table includes columns for 'Parameter', 'Value', 'Uncert. Ext.', 'units', 'Driver', 'Standard uncertainty', 'units', 'Series', 'Coef.', 'units', and 'Notes'. Below the table, there is a section for 'Effect of temperature' with a table showing 'Current force range' and 'Temperature error at the calibration' for values 1 through 5.


Parameter	Value	Uncert. Ext.	units	Driver	Standard uncertainty	units	Series	Coef.	units	Notes			
Accuracy of reference LC	B	0	R	1.14	MN	1.73305	0.5980E-01	MN	1	MN/M	0.5980E-01	100%	infinity
Resolution of force meas.	B	0	R	2.77	MN	1.73305	1.5980E-00	MN	1	MN/M	1.5980E-00	100%	infinity
Force	B	300	N	0.10	MN	1.73305	5.7730E-02	MN	1	MN/M	5.7730E-02	100%	infinity
Reproducibility	A	300	N	0.84	MN	1.73305	4.8778E-01	MN	1	MN/M	4.8778E-01	100%	6
Effect of temperature	B	0	R	8.06	MN	1.73305	2.0980E-02	MN	1	MN/M	2.0980E-02	100%	infinity

The result of the measurement consists of:


- 1: 300 MN
- 2: -0.00%
- 3: 0.05%
- 4: 0.14%
- 5: 0.8 N
- 6: 0.02%

$$v_{\text{eff}} = \frac{\sum_{i=1}^N u_i^4(y)}{\sum_{i=1}^N v_i}$$

The detail



The process



Why do we test new ropes?

Theory

$$UTS_{rope} \neq \sum UTS_{t_wires}$$




vs.

Practice

Rope strength depends on:


- Material properties of wires
- Wire uniformity
- Rope construction

Component Parts of Wire Rope



Why do we test ropes in service?

- Rope condition will deteriorate with use
- Rate of deterioration cannot be predicted
- Conveyance is held by one rope only
- MHSA Reg 16.41 – mandatory tensile test and
- SANS 10293 CoP – discard criteria ($UTS_{used} \geq 0.9 UTS_{new}$)



our future through science

High force, high stakes environment

- MFL Tensile tester
 - 15 000 kN capacity testing from 750 kN (5% of FS)
 - 2 m stroke length, 8 m daylight
- Laboratory space 1200m², not climate controlled
- Calibration duration, cost
- No CRM, no PT schemes
- No re-test of the same UUT



Enter uncertainty...

- ? Instrumentation
- ? Environment e.g. temperature and humidity
- ? Variability
- ? Resolution of force measurement
- ? Electromagnetic noise
- ? Test method
- ? Personnel



The estimation process


The GUM method

- Model the measurement
- Identify and quantify all sources of error
- Evaluate all sources of error to obtain the uncertainty contribution
- Combine uncertainty contributions
- Calculate the degrees of freedom for the combined uncertainty
- Choose level of confidence
- Calculate the expanded uncertainty, U
- Report the result

The detail – Mathematical model


- $\sigma = E\varepsilon$
- $\varepsilon = \frac{1}{GF} \Delta R / R$
- $\varepsilon = \frac{-2V_r}{GF[(\nu+1)-V_r(\nu-1)]}$

Where $V_r = \frac{V_{ch_strained} - V_{ch_unstrained}}{V_{ex}}$



or

$$F_{apply} = F_{record}$$



Sources of error

- Accuracy of machine force measurement system
- Accuracy of reference load cell – uncertainty quoted by calibration lab
- Resolution of force measurement
- Repeatability
- Reproducibility
- Effect of temperature




our future through science

Uncertainty budget

	Symbol	Description	Type	Exp Val	Distr	Uncert. Est.	units	Divisor	Standard uncertainty	units	Sens. Coeff.	units	Uncertainty contributor	Reliability	Deg of Freedom	Significance
1									$u(x_i)$		c_i		$u(y_i)$		v_i	
2																
3																
4	F_{accLC}	Accuracy of LC + amp	B	300	R	1.14	kN	1.73205	6.5886E-01	kN	1	KN/kN	6.5886E-01	100%	infinite	12.54%
5	F_{acc_refLC}	Accuracy of reference LC	B	0	R	2.77	kN	1.73205	1.5987E+00	kN	1	KN/kN	1.5987E+00	100%	infinite	73.83%
6	F_{res}	Resolution of force meas.	B	0	R	0.10	kN	1.73205	5.7735E-02	kN	1	KN/kN	5.7735E-02	100%	infinite	0.10%
7	F_{repr}	Reproducibility	A	300	N	0.67	kN	1	6.7171E-01	kN	1	KN/kN	6.7171E-01	100%	5	13.03%
8	F_{var}	Repeatability	A	300	N	0.13	kN	1	1.2849E-01	kN	1	KN/kN	1.2849E-01	100%	3	0.48%
9	F_{temp}	Effect of temperature	B	0	R	0.05	kN	1.73205	2.5981E-02	kN	1	KN/kN	2.5981E-02	100%	infinite	0.02%
10																
11													$u_c(y)$	1.86054E+00		
12													v_{eff}	293.65		
13	Working point in force range:			1	300	kN							Level of Confidence	95.45%		
14	Mean indication error at the calibration point			2	-0.38%								Coverage factor	2.009		
15	Ref LC uncertainty			3	0.92%								U	3.7378153	kN	
16	Repeatability			4	Lookup table								or	1.246%		
17	Reproducibility			5	Lookup table											
18	Temp effect on force measurement			6	0.02%											
19																
20	Table 1 - Summary of estimated uncertainty												The result of the measurement was found to be (300.0 ± 3.8) kN with a coverage factor of 2.009 at a level of confidence of 95.45%			

Temperature effect

Breaking force (kN)




Summer: +0.3%
Winter: -0.3%

$\Delta T = 20^\circ\text{C}$ (12 months to date)

Seasonal diff from mean
rope BF

Difference from mean BF (n = 12)



The outcome – Example 1 at 852 kN working point

	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	
1	Symbol	Description	Type	Expected Value	Distribution	Uncertainty Estimate	units	Divisor	Standard uncertainty	units	Sensitivity coefficient	units	Uncertainty contributor	Reliability	Degrees of Freedom	Significance	
2									$u(x_i)$		c_i		$u(y_i)$		v_i		
4	F_{accLC}	Accuracy of LC + amp	B	0	R	2.6	kN	1.7321	1.4757E+00	kN	1	kN/kN	1.4757E+00	100%	infinite	30.28%	
5	F_{acc_refLC}	Accuracy of reference LC	B	0	R	2.0	kN	1.7321	1.1806E+00	kN	1	kN/kN	1.1806E+00	100%	infinite	19.38%	
6	F_{res}	Resolution of force meas.	B	0	R	0.1	kN	1.7321	5.7735E-02	kN	1	kN/kN	5.7735E-02	100%	infinite	0.05%	
7	F_{var}	Variability in force meas.	A	852	N	1.2	kN	1	1.2000E+00	kN	1	kN/kN	1.2000E+00		5	20.02%	
8	F_{temp}	Effect of temperature	B	0	R	2.6	kN	1.7321	1.4757E+00	kN	1	kN/kN	1.4757E+00	100%	infinite	30.28%	
9																	
10											$u_c(y)$		2.68188E+00				
11											V_{eff}		124.73				
12													Level of Confidence	95.45%			
13													Coverage factor	2.020			
14													U	5.4174024	kN	or	0.636%
15																	
16													The result of the measurement was found to be (852.0 ± 5.5) kN with a coverage factor of 2.02 at a level of confidence of 95.45%				

The outcome – Example 2 at 1965 kN working point

	A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P
1	Symbol	Description	Type	Expected Value	Distribution	Uncertainty Estimate	units	Divisor	Standard uncertainty	units	Sensitivity coefficient	units	Uncertainty contributor	Reliability	Degrees of Freedom	Significance
2									$u(x_i)$		c_i		$u(y_i)$		v_i	
3																
4	F_{accLC}	Accuracy of LC + amp	B	0	R	3.9	kN	1.7321	2.2690E+00	kN	1	kN/kN	2.2690E+00	100%	infinite	18.08%
5	F_{acc_refLC}	Accuracy of reference LC	B	0	R	4.7	kN	1.7321	2.7228E+00	kN	1	kN/kN	2.7228E+00	100%	infinite	26.03%
6	F_{res}	Resolution of force meas.	B	0	R	1.0	kN	1.7321	5.7735E-01	kN	1	kN/kN	5.7735E-01	100%	infinite	1.17%
7	F_{var}	Variability in force meas.	A	1965	N	2.0	kN	1	2.0000E+00	kN	1	kN/kN	2.0000E+00		5	14.05%
8	F_{temp}	Effect of temperature	B	0	R	5.9	kN	1.7321	3.4035E+00	kN	1	kN/kN	3.4035E+00	100%	infinite	40.67%
9																
10											$u_c(y)$		5.33656E+00			
11											v_{eff}		253.45			
12											Level of Confidence		95.45%			
13											Coverage factor		2.010			
14											U		10.7264833	kN	or	0.546%
15																
16											The result of the measurement was found to be (1 965 ± 11) kN with a coverage factor of 2.01 at a level of confidence of 95.45%					

Room for improvement

- Temperature effect
 - Control lab temperature
 - Condition specimens prior to testing
 - Apply a temperature correction to test result
- Accuracy of machine force measurement
 - Apply correction
 - Adjust machine force calculation polynomial coefficients
- Uncertainty of reference load cell
 - Load cell?
 - Calibration laboratory



our future through science



Thank you.